

# Quality And Accreditation Institute

## Centre for Laboratory Accreditation

### Scope of Accreditation

#### **MEDICAL LABORATORY SERVICES (RAJIV GANDHI CENTRE FOR BIOTECHNOLOGY)**

District Hospital,  
Nedumangad, Thiruvananthapuram - 695541  
Kerala, India

QAI/CLA/ML/2019/0011

Valid from : 4 December 2019  
Valid until : 3 December 2021

#### **Accreditation Standard: ISO 15189:2012**

<b>Clinical Biochemistry</b>					
Sl. No.	Type of Samples Examined/ Tested	Specific tests/ Examination Performed	Standard (Method), Principle / Methodology or Technique used	Range of testing/ Limit of Detection	%CV
1.	Serum	Uric Acid	Uricase Method	0.20 - 25 mg/dl	Level 1 - 1.87 Level 2 - 1.99
2.	Serum	Glucose	Hexokinase Method	2 - 750 mg/dl	Level 1 - 2.71 Level 2 - 2.99
3.	Serum	Cholesterol	Cholesterol Oxidase, Esterase, Peroxidase Method	3.86 - 800 mg/dl	Level 1 - 2.93 Level 2 - 3.74



A handwritten signature in black ink, appearing to read 'Dr. Bhupendra Kumar Rana', is written over a horizontal line.