

Quality And Accreditation Institute

Centre for International Accreditation



Certificate of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test Laboratory)

*5th, Floor, Business Bay, Jito, Mumbai, Naka, Nashik- 422001,
Maharashtra, India*

*has been assessed and accredited in accordance with the Standard
ISO/IEC 17025:2017*

*"General Requirements for the Competence of Testing and Calibration Laboratories"
In the field of*

Testing

*This certificate remains valid for the Scope of Accreditation as specified in
the annexure subject to continued compliance with the above standard &
any other requirements specified by QAI.*



QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Dr. Bhupendra Kumar Rana
Chief Executive Officer

Prof. Vikram Kumar
Chair, CIA



Quality And Accreditation Institute

Centre for Laboratory Accreditation



Scope of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test laboratory)

5th, Floor, Business Bay, Jito, Mumbai Naka,
Nashik- 422001, Maharashtra, India

QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Accreditation Standard: ISO/IEC 17025:2017

Diagnostic Radiology QA Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Computed Tomography		
1.	Computed Tomography	Accuracy of Operating Potential (kV)	CT Equipment
2.	Computed Tomography	Accuracy of Timer (s)	CT Equipment
3.	Computed Tomography	High Contrast Resolution	CT Equipment
4.	Computed Tomography	Low Contrast Resolution	CT Equipment
5.	Computed Tomography	Linearity of Radiation Output (mA/MAs Linearity)	CT Equipment
6.	Computed Tomography	Measurement of Radiation Protection Survey	CT Equipment
7.	Computed Tomography	Radiation Dose Index CTDI	CT Equipment
8.	Computed Tomography	Radiation Leakage Level from X-Ray Tube Housing	CT Equipment
9.	Computed Tomography	Radiation Profile width/ Slice Thickness	CT Equipment
10.	Computed Tomography	Reproducibility of Output COV	CT Equipment

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



1 of 6

Quality And Accreditation Institute

Centre for Laboratory Accreditation



Scope of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test laboratory)

5th, Floor, Business Bay, Jito, Mumbai Naka,
Nashik- 422001, Maharashtra, India

QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Accreditation Standard: ISO/IEC 17025:2017

Diagnostic Radiology QA Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Computed Tomography		
11.	Computed Tomography	Total Filtration	CT Equipment
12.	Computed Tomography	Mechanical Test (Gantry Tilt)	CT Equipment
13.	Computed Tomography	Mechanical Test (Alignment of Table to Gantry)	CT Equipment
14.	Computed Tomography	Mechanical Test (Table Indexing Accuracy)	CT Equipment



Quality And Accreditation Institute

Centre for Laboratory Accreditation



Scope of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test laboratory)

5th, Floor, Business Bay, Jito, Mumbai Naka,
Nashik- 422001, Maharashtra, India

QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Accreditation Standard: ISO/IEC 17025:2017

Diagnostic Radiology QA Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Radiography (X-ray -Fixed Portable, Mobile)		
1.	Radiography (X-ray -Fixed Portable, Mobile)	Accuracy of Irradiation Time(s)	X-Ray (Fixed, Portable and Mobile)
2.	Radiography (X-ray -Fixed Portable, Mobile)	Accuracy of Operating Potential (kV)	X-Ray (Fixed, Portable and Mobile)
3.	Radiography (X-ray -Fixed Portable, Mobile)	Central Beam Alignment	X-Ray (Fixed, Portable and Mobile)
4.	Radiography (X-ray -Fixed Portable, Mobile)	Congruence of Radiation and Optical Field	X-Ray (Fixed, Portable and Mobile)
5.	Radiography (X-ray -Fixed Portable, Mobile)	Consistency of Radiation Output (CoV)	X-Ray (Fixed, Portable and Mobile)
6.	Radiography (X-ray -Fixed Portable, Mobile)	Effective Focal Spot Measurement FFD= 60 cm	X-Ray (Fixed, Portable and Mobile)
7.	Radiography (X-ray -Fixed Portable, Mobile)	Linearity of mA/MAs Loading Stations	X-Ray (Fixed, Portable and Mobile)
8.	Radiography (X-ray -Fixed Portable, Mobile, BMD)	Measurement of Radiation Protection Survey	X-Ray (Fixed, Portable and Mobile)

This is annexure to 'Certificate of Accreditation' and does not require any signature.

www.qai.org.in



Quality And Accreditation Institute

Centre for Laboratory Accreditation



Scope of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test laboratory)

5th, Floor, Business Bay, Jito, Mumbai Naka,
Nashik- 422001, Maharashtra, India

QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Accreditation Standard: ISO/IEC 17025:2017

Diagnostic Radiology QA Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Radiography (X-ray -Fixed Portable, Mobile)		
9.	Radiography (X-ray -Fixed Portable, Mobile, BMD)	Radiation Leakage Level at 1m from Tube Housing and Collimator	X-Ray (Fixed, Portable and Mobile)
10.	Radiography (X-ray -Fixed Portable, Mobile)	Total Filtration	X-Ray (Fixed, Portable and Mobile)



Quality And Accreditation Institute

Centre for Laboratory Accreditation



Scope of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test laboratory)

5th, Floor, Business Bay, Jito, Mumbai Naka,
Nashik- 422001, Maharashtra, India

QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Accreditation Standard: ISO/IEC 17025:2017

Diagnostic Radiology QA Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Interventional Radiology & Fluoroscopy		
1.	Interventional Radiology & Fluoroscopy	Accuracy of Irradiation Time (s)	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
2.	Interventional Radiology & Fluoroscopy	Accuracy of Operating Potential (kV)	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
3.	Interventional Radiology & Fluoroscopy	Consistency of Radiation Output (CoV)	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
4.	Interventional Radiology & Fluoroscopy	Linearity of mA/MAs Loading Stations	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
5.	Interventional Radiology & Fluoroscopy	Low Contrast Resolution	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
6.	Interventional Radiology & Fluoroscopy	Measurement of Radiation Protection Survey	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
7.	Interventional Radiology & Fluoroscopy	Total Filtration	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine



Quality And Accreditation Institute

Centre for Laboratory Accreditation



Scope of Accreditation

Sai Imaging Healthcare Private Limited
(QA Test laboratory)

5th, Floor, Business Bay, Jito, Mumbai Naka,
Nashik- 422001, Maharashtra, India

QAI/CIA/TL/2026/0183

Valid from: 20 August 2026

Valid until: 19 August 2028

Accreditation Standard: ISO/IEC 17025:2017

Diagnostic Radiology QA Testing			
Sl. No.	Product(s)/Material of Test	Specific Tests Performed	Test Method
	Interventional Radiology & Fluoroscopy		
8.	Interventional Radiology & Fluoroscopy	Exposure Rate at Tabletop	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
9.	Interventional Radiology & Fluoroscopy	High Contrast Resolution	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine
10.	Interventional Radiology & Fluoroscopy	Radiation Leakage Level at 1m from tube Housing and Collimator	C-ARM, O-ARM, Cath Lab, Radiography and Fluoroscopy Machine

