Quality And Accreditation Institute

Centre for International Accreditation

Certificate of Accreditation

Jodhpur Calibration Laboratory
Indian Oil Corporation Limited (Marketing Division)

Jodhpur Terminal, Near Salawas Railway Station
Jodhpur-342013, Rajasthan, India

has been assessed and accredited in accordance with the Standard

ISO/IEC 17025:2017

"General Requirements for the Competence of Testing and Calibration Laboratories"
In the field of

Calibration

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued compliance to the above standard & any other requirements specified by QAI.

QAI/CIA/CL/2026/0013

Valid from: **26 October 2026**

Valid until: **25 October 2028**

Dr. Bhupendra Kumar Rana
Chief Executive Officer

Prof. Vikram Kumar
Chair, CIA

User is advised to verify the current certificate and scope of accreditation by visiting our website: www.qai.org.in

Quality And Accreditation Institute
Centre for International Accreditation

Scope of Accreditation

Jodhpur Calibration Laboratory

Indian Oil Corporation Limited (Marketing Division)

Jodhpur Terminal, Near Salawas Railway Station,
Jodhpur-342013, Rajasthan, India

QAI/CIA/CL/2026/0013

Valid from: 26 October 2026
Valid until: 25 October 2028

Accreditation Standard: ISO/IEC 17025:2017

Mechanical Discipline- Density & Viscosity				
Sl. No.	**Measurand or Reference Material/ Type of Instrument or Material to be Calibrated or Measured/ Quantity Measured/ Instrument**	**Calibration or Measurement Method or Procedure**	**Measurement Range and Additional Parameters Where Applicable (Range and Frequency)**	***Calibration and Measurement Capability (CMC) (±)**
At Permanent Laboratory				
1.	Density Hydrometer	Using Standard Hydrometer and density Liquids by Comparison Method as per IS 3104 (Part-II)-1982	0.700-0.750 g/cm^3	±0.0009 g/cm^3
2.	Density Hydrometer	Using Standard Density Hydrometer and density Liquids by Comparison Method as per IS 3104 (Part-II)-1982	0.750-0.800 g/cm^3	±0.0009 g/cm^3
3.	Density Hydrometer	Using Standard Density Hydrometer and density Liquids by Comparison Method as per IS 3104 (Part-II)-1982	0.800-0.850 g/cm^3	±0.0009 g/cm^3

Scope of Accreditation

Jodhpur Calibration Laboratory
Indian Oil Corporation Limited (Marketing Division)

Jodhpur Terminal, Near Salawas Railway Station,
Jodhpur-342013, Rajasthan, India

QAI/CIA/CL/2026/0013

Valid from: 26 October 2026
Valid until: 25 October 2028

Accreditation Standard: ISO/IEC 17025:2017

Thermal Discipline- Temperature				
Sl. No.	**Measurand or Reference Material/ Type of Instrument or Material to be Calibrated or Measured/ Quantity Measured/ Instrument**	**Calibration or Measurement Method or Procedure**	**Measurement Range and Additional Parameters Where Applicable (Range and Frequency)**	***Calibration and Measurement Capability (CMC) (±)**
At Permanent Laboratory				
1.	Liquid in Glass Thermometer	Using RTD with Temperature Indicator & Liquid Bath by Comparison Method as per IS 6274:2026	0 to 30°C	±0.40°C
2.	Liquid in Glass Thermometer	Using RTD with Temperature Indicator & Liquid Bath by Comparison Method as per IS 6274:2026	30 to 80°C	±0.40°C

*** CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.**