Quality And Accreditation Institute

Centre for International Accreditation

Certificate of Accreditation

Neutech Automation

2/45, SFS, Agarwal Farm, Mansarovar,
Jaipur-302020, Rajasthan, India

has been assessed and accredited in accordance with the Standard

ISO/IEC 17025:2017

"General Requirements for the Competence of Testing and Calibration Laboratories"
In the field of

Calibration

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued compliance to the above standard & any other requirements specified by QAI.

QAI/CIA/CL/2026/0014

Valid from: ***13 November 2026***

Valid until: ***12 November 2028***

Dr. Bhupendra Kumar Rana
Chief Executive Officer

Prof. Vikram Kumar
Chair, CIA

User is advised to verify the current certificate and scope of accreditation by visiting our website: www.qai.org.in

Quality And Accreditation Institute
Centre for International Accreditation

Scope of Accreditation

Neutech Automation

2/45, SFS, Agarwal Farm, Mansarovar,
Jaipur-302020, Rajasthan, India

QAI/CIA/CL/2026/0014

Valid from: 13 November 2026
Valid until: 12 November 2028

Accreditation Standard: ISO/IEC 17025:2017

Thermal Discipline- Temperature				
Sl. No.	**Measurand or Reference Material/ Type of Instrument or Material to be Calibrated or Measured/ Quantity Measured/ Instrument**	**Calibration or Measurement Method or Procedure**	**Measurement Range and Additional Parameters Where Applicable (Range and Frequency)**	***Calibration and Measurement Capability (CMC) (±)**
At Permanent and At-Site Facility				
1.	RTD/ Thermocouple (J, K, R, S & N Type)	Comparison Method Procedure No. NA-QMS-P-07	200 °C to 1200 °C	2.62 °C
2.	Temperature Controllers/ RTD (J, K, R, S & N Type)	Comparison Method Procedure No. NA-QMS-P-06	200 °C to 1700 °C	0.68 °C

*** CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.**